

Prognost XP

**mobile
Patient Table
with floating tabletop**

Model/ID: 7036-0-17xx

Instruction for Use

Ident. Nr. 5031-0-0002



ALL SHEETS OF THIS DOCUMENT CONTAIN PROPRIETARY AND CONFIDENTIAL INFORMATION OF **PROTEC GmbH & Co. KG** AND IS INTENDED FOR EXCLUSIVE USE BY CURRENT **PROTEC GmbH & Co. KG** CUSTOMERS. COPYING, DISCLOSURE TO OTHERS OR OTHER USE IS PROHIBITED WITHOUT THE EXPRESS WRITTEN AUTHORIZATION OF **PROTEC GmbH & Co. KG**'S LAW DEPARTMENT. REPORT ANY VIOLATIONS OF THIS REQUIREMENT TO **PROTEC GmbH**.

© 2015 PROTEC GmbH & Co.KG, Oberstenfeld

This document is prepared and distributed by Publications Department.
Send inquiries regarding this document to the following address:

PROTEC GmbH & Co.KG

In den Dorfwiesen 14 | 71720 Oberstenfeld
Deutschland

Tel: (+ 49) 7062 – 92 55 0

Fax: (+ 49) 7062 – 22 68 5

E-Mail: protec@protec-med.com

Internet: www.protec-med.com

Content

	Page
Content	i
NOTE	a
Document Effectivity	a
Mechanical - Electrical Warning	c
Radiation Warning	c
To the User	c
Improvement Recommendations	c
1 Equipment Description	1
1.1 Introduction	1
1.2 Intended Purpose	1
1.3 Declaration of Conformity	1
1.4 Description	2
1.5 Unit versions	2
1.6 Optional Equipment	2
1.7 Nameplate	2
2 Control Elements and Indicators	3
2.1 Brakes	3
2.2 Brake for Tabletop	3
2.3 Hand grips (optional)	4
2.4 Deflector (optional)	4
2.5 Side edge guard (optional)	4
2.6 Bumper profile (optional)	4
2.7 Protection film (optional)	4
2.8 Center stop (optional)	4
2.9 Labels on head and foot end of table top	4
2.10 Labels on table top	5
2.11 Label on front side frame	5
3 Operating Instructions	6
3.1 Safety Aspects	6
3.1.1 Requirements for Operation	6
3.1.2 Users	6
3.1.3 Radiation Protection	6
3.1.4 Interferences to other devices	6
3.1.5 Warnings	7
3.1.6 Pinch points and possible collisions	7
3.2 Adjustment of Exposure Position	8
3.2.1 Exposures with PROGNOST XP	8
4 Operator Maintenance	9
4.1 Introduction	9
4.1.1 Operating	9
4.2 Safety Information	9
4.3 Technical Safety Information	9
4.4 Technical safety checks	10
4.5 Maintenance schedule	10
4.5.1 User's maintenance schedule	10
4.5.1.1 Cleaning	10
4.5.1.2 Disinfection	10
4.5.2 User's daily maintenance before and during operation	10
4.5.3 Monthly Checks	11
4.5.3.1 Quality Control	11
4.5.4 Maintenance	11

4.5.5	Duration of Lifetime of the Product.....	11
4.5.6	Disposal Remarks.....	11
5	Combination with other equipment	11
6	Technical Data.....	12
6.1	Measurements.....	12
6.2	Attenuation equivalent of Tabletop	12
6.3	Product Life Time.....	12
6.4	Environmental Conditions	12
6.4.1	Operating Environment.....	12
6.4.2	Transport and stock environment	12
7	Description of Symbols, Labels and Abbreviations	13
7.1	Symbols	13
7.2	Labels	13
7.3	Abbreviations	14

NOTE

The information contained in this document conforms with the configuration of the equipment as of the date of manufacture. Revisions to the equipment subsequent to the date of manufacture will be addressed in service updates distributed to the PROTEC GmbH & Co. KG Technical Service Organization.

Document Effectivity

Revision No.	Date	List of effective pages	Comments
	27/07/2007		Original Issue
01	23/01/2008	all	UL Label changed, accessories supplemented
02	23/11/2009		Change from GmbH & Co. KG & Co. KG to GmbH
03	07/04/2011	all	Modified address
04	15/08/2011	All	Protec Label
05	01/12/2011		New Label
06	03/05/2012		New nameplate
07	02/08/2013		Changed standard / new label
08	07/01/2014		Technical Safety check interval 2 years
09	10/02/2015		Changing Labels
10	27/03/2018	Title	Corrected Modelnumber
11	2020-11-24	Front page	Model ID revised

Mechanical - Electrical Warning

All of the **movable assemblies and parts of this equipment** should be operated with care and routinely inspected in accordance with the manufacturer's recommendations contained in the equipment accompanying documents. Maintenance and service is only to be performed by personnel authorized by **PROTEC GmbH & Co.KG**

Live electrical terminals are deadly.

Do not remove flexible high tension cables from X-ray tube housing or high tension generator and/or access covers from X-ray generator.

For all components of the equipment protective earthing means must be provided in compliance with the national regulations.

Failure to comply with the foregoing may result in serious or fatal bodily injuries to the operator or those in the area.

Radiation Warning

The component of the equipment described within this Document is part of a system for the intended generation of X-rays for medical diagnosis.

X-rays generate a potential risk for both patients and operators.

For this reason the application of X-rays for a given medical purpose must aim at the minimization of radiation exposition to any persons.

Those persons responsible for the application must have the specific knowledge according to legal requirements and regulations and must establish safe exposure procedures for these kind of systems. Those persons responsible for the planning and installation of this equipment must observe the national regulations.

To the User

The user of this Document is directed to read and carefully review the instructions, warnings and cautions contained herein prior to beginning operation, installation or service activities. While you may have previously operated equipment similar to that described in this Document, changes in design, manufacture or procedure may have occurred which significantly affect the present operation.

The installation and service of equipment described herein is to be performed by authorized, qualified **PROTEC GmbH & Co. KG** personnel. Assemblers and other personnel not employed by nor directly affiliated with **PROTEC GmbH & Co. KG** technical services are directed to contact the local **PROTEC GmbH & Co. KG** office before attempting installation or service procedures.

Improvement Recommendations

Users of this Document are encouraged to report errors, omissions and improvement recommendations to **PROTEC GmbH & Co.KG**

1 Equipment Description

1.1 Introduction

This „Instruction for Use“ describes the special, characteristic features and the correct operational instructions which are necessary for the efficient and effective use of the Prognost XP.

It is suggested that you review the operation instructions, the safety notes and the controls described in this “Instruction for Use” before using the Prognost XP. Each control device and each display is described in order to make you acquaint with its function.

1.2 Intended Purpose

The PROGNOST XP is a patient table for examinations with X-ray units (e.g. type of L/U-arm units) for general radiography in diagnostic human medicine use. It has to be used as a movable patient table in medically used rooms only.

The PROGNOST XP is not suitable for patient transportation.

The Prognost XP must be operated by qualified users trained in the use of human diagnostic X-ray equipment in conformity with the respective national regulations.
An introduction of the product is necessary for the user.

1.3 Declaration of Conformity



This product is in conformity with the requirements of the European Community Medical Device Directive 93/42/EEC from 06/14/1993 including all current revision standards

Upon request the declaration of conformity is available from:

PROTEC GmbH & Co.KG

In den Dorfwiesen 14 | 71720 Oberstenfeld
Deutschland

Tel: (+ 49) 7062 – 92 55 0

Fax: (+ 49) 7062 – 22 68 5

E-mail: protec@protec-med.com

Internet: www.protec-med.com

1.4 Description

- large range of application
- high reliability
- easy movement
- floating tabletop

1.5 Unit versions

Composite fiber frame (choose one)

- **PROGNOST XP**
Frame color RAL 9003 Model/ID: 7036-0-17xx

Table top versions (choose one)

Model ID	Material	L cm (inch)	W cm (inch)	Table top color
7301-0-5900	Carbon	200 (78.74)	75,5 (29.72)	white
7301-0-2200	Carbon	226 (88.97)	75,5 (29.72)	white
7301-0-6000	Composite fiber	200 (78.74)	75,5 (29.72)	white
7301-0-6010	Composite fiber	226 (88.97)	75,5 (29.72)	white
7301-0-6020	Composite fiber	200 (78.74)	65,5 (25.78)	white

1.6 Optional Equipment

- **Handle**, mounted at rear side of the table top in order to ease the patient's getting on and off.
- **Hand grips**, mounted at front side of the table in order to ease movement of the table or of the tabletop.
- **Deflector**, for easier positioning of the table above digital detector
- **Side edge guard**, for protection of the detector
- **Bumper profile**, for rear accessory rail of table top
- **Protection film**, for internal covers
- **Center stop**, for a pinpoint positioning

1.7 Nameplate

For nameplate location refer to Figure 2-1/item 4.

2 Control Elements and Indicators

2.1 Brakes

In order to fix the patient table in all directions, it is fitted with 2 brake wheels at the front, which have to be fixed with a fastener (Figure 2-1/1) on the left or right side before a patient gets on or off the tabletop.




Figure 2-1

1. Brake lever
2. Foot pedal bar for table top brakes
3. Table top
4. Nameplate, UL label, FDA label
5. Rear side handle (Optional)
6. Front side hand grips (Optional)
7. Side edge guard for protection of the detector
8. Deflector for easier positioning of the table above digital detector
9. Bumper profile for rear accessory rail of table top
10. Protection film for internal covers

2.2 Brake for Tabletop

Depress the brake foot-pedal bar (Figure 2-1/2) for manual movement of the tabletop.

2.3 Hand grips (optional)

As options a long handle at rear side of the table top (Fig. 2-1/5) and short hand grips (Fig. 2-1/6) at front side of the table top are available. The hand grips and handle can be removed with tools only. The long handle can be used for patient to easier getting ON and OFF the table top. The short hand grips allow more convenient movement of table and table top.

2.4 Deflector (optional)

The deflectors (Fig. 2-1/8) at the transverse carriage avoid contact of the u-rail of transverse carriage with detector housing and consequently facilitate the positioning of the table.

2.5 Side edge guard (optional)

The side edge guards (Fig. 2-1/7) avoid direct contact of detector housing with the z-angle.

2.6 Bumper profile (optional)

The protection profile (Fig. 2-1/9) avoids a direct contact of rear accessory rail of the table with the stand or detector support.

2.7 Protection film (optional)


The protection film (Fig. 2-1/10) on internal covers increases the slippage at contact with detector housing.

2.8 Center stop (optional)

The center stop (not shown) for cross traverse of table top improves alignment of the table to x-ray tube assembly.

2.9 Labels on head and foot end of table top


Tabletop made of carbon



Tabletop Composite fiber 200cm



Tabletop Composite fiber 226cm



Attention

When the tabletop is completely moved forwards, backwards, to head or foot end, the maximum permissible load on the outside edges is:

Otherwise whole table may tilt or can tip over!

For heavier patients heavier move table top to approx. center position before they get ON and OFF the table top

Before patients get ON or OFF the table, the PROGNOST XP brake wheels must be fixed in all directions with brake lever (Figure 2-1/1)

2.10 Labels on table top



Caution: Pay attention to squeeze possibilities of fingers or hands during movement of table top and x-ray unit.



230kg
506lb

Max. allowed patient weight on table top (distributed load)
Composite fiber tabletop



250kg
550lb

Max. allowed patient weight on table top (distrubuted load)
Carbon tabletop

2.11 Label on front side frame



3 Operating Instructions

3.1 Safety Aspects

3.1.1 Requirements for Operation

The mobile patient tables are a component of the general x-ray systems as e.g. L/U-arm units.

The X-ray system with the mobile patient table PROGNOT XP must be completely installed and officially handed over by **PROTEC** personnel or by service personnel, authorized by **PROTEC**, before it can be used by the customer.

The national regulations concerning announcement and release of new installed X-ray equipment and monitoring of the operating system must be met.

Maintenance has to be carried out according to the instructions in chapter 4.

3.1.2 Users

The system must be operated by qualified users, trained in the use of human diagnostic X-ray equipment in conformity with the respective national regulations.

A product introduction for the user is necessary.

3.1.3 Radiation Protection

X-rays represent a hazard to both patients and operator if the rules for the use of such systems are not observed.

For this reason the principles of radiation protection must have the highest priority and they shall be observed at all times.

- **Keep distance from the radiation source**

The dose declines with the square of the distance from a (point-like) radiation source, i.e. doubling the distance reduces the dose to one quarter, tripling the distance reduces the dose to one ninth, etc.

- **Keep exposure periods short**

The dose rises linear to the exposure time, i.e. halving the exposure period will halve the dose (this is applicable particularly for fluoroscopy; when making X-ray films the current time product (mAs value) is prescribed in the most cases).

- **Use shielding and protective clothing**

The protective factor rises exponentially with the thickness of the shielding. This means that two half-value layers will reduce (homogeneous) radiation to 1/4, three half-value layers to 1/8, and 10 half-value layers will reduce the radiation to less than 1/1000 of the original value.

- **Never reach into the direct X-ray beam**

The dose in the direct, non-attenuated X-ray beam is some 100 times higher than the scattered radiation.

- **Personal dose meters**

During work with X-ray use corresponding personal dose meters for measurement of the accumulated dose.

When taking X-rays near the reproductive organs, pay attention to using the best possible protection (testicle shielding cup or lead apron).




















The operator shall always stand behind a shielding panel or partition when taking exposures.

Persons who must be near the patient during fluoroscopy shall wear protective clothing (a lead apron, for instance). The same applies for maintenance and repair work.

3.1.4 Interferences to other devices

There are no interferences to other devices known.

3.1.5 Warnings

	Keep enough distance between moving parts during movements.										
<table border="1"> <tr> <td data-bbox="245 405 331 488">  </td> <td data-bbox="331 405 523 488">  </td> <td data-bbox="523 405 850 488"> 100kg (220 lb) </td> </tr> <tr> <td data-bbox="245 533 331 616">  </td> <td data-bbox="331 533 523 616">  </td> <td data-bbox="523 533 850 616"> 75kg (165 lb) </td> </tr> <tr> <td data-bbox="245 660 331 743">  </td> <td data-bbox="331 660 523 743">  </td> <td data-bbox="523 660 850 743"> 60kg (132 lb) </td> </tr> </table>			100kg (220 lb)			75kg (165 lb)			60kg (132 lb)	When tabletop is completely shifted forwards, backwards, to the left or right, the maximum permissible load on the outside edges is as indicated on head or foot end of table top. Otherwise whole table may tilt!	
		100kg (220 lb)									
		75kg (165 lb)									
		60kg (132 lb)									


References regarding the maximum permissible load of the table, attached on the tabletop of the **PROGNOST XP**, have to be observed especially when patients get on or off the table. For heavier patients move tabletop to center position above the table base frame!

3.1.6 Pinch points and possible collisions

CAUTION!

Hands and fingers may be squeezed at the shown arrow positions.

Be sure that during movement of the table, tabletop and x-ray unit neither the patient or personal are within an uncontrolled movement area.



CAUTION!

Take care that at equipment, which can be raised or lowered or which can be moved in different directions yourself or other persons are not within the movement area. Remove objects, as e.g. chairs, out of the collision area, before using the table.

Pay attention that careless movements of the PROGNOST XP table and table top may injure patients, users or other persons and may cause damages to the X-ray tube assembly or the X-ray image receptor. Non-observance can cause bodily injuries and damages of the x-ray unit.

3.2 Adjustment of Exposure Position

Before positioning of table with patient, bring x-ray unit into right exposure position.

3.2.1 Exposures with PROGNOST XP

The following handling is necessary in order to avoid a collision between PROGNOST XP, X-ray image receptor and other furnishing

- Bring X-ray unit as e.g. a L/U-arm in 0-degree position.
- Adjust height of X-ray image receptor to table height
- Move table to check there is no collision with image receptor
- Move table into a position for convenient patient get on the table top.
- Actuate brake lever for front brake wheels and check that wheels are effectively fixed.

Attention:

**The PROGNOST XP is not suitable for patient transportation.
It only can be used as a support for the patient during examination.
Patients are only allowed to get on or off the tabletop if the brakes are fixed (Chap. 3.1.7) and under supervision and assistance from the examining personnel, as otherwise the whole table may tilt.**

- Assist patient to get on the table top
- Place patient for examination. Use where required (e.g. at open wounds) suitable clothes to cover the table top surface.



CAUTION:

Caution for squeeze points at table borders and below and above table top

At horizontal movement of the table and the table top extremities can be squeezed between table top and fixed obstacles (wall, columns, x-ray unit).

Therefore:

Watch during movement of the PROGNOST XP that patient and personal are not within area of movement direction.

Especially take care that no patient extremities protrude over the table top borders

- Release brake lever and move PROGNOST XP into exposure position.
- Actuate brake lever.
- Depress foot pedal to release table top brake and move table top with patient to exact exposure position and release foot pedal to fix table top brakes.
- Follow instructions of the x-ray unit to take an exposure.
- After the examination release wheel brakes and move table to a position for convenient patient to get off the table top.
- Actuate brake lever and assist patient to get off the table top.

4 Operator Maintenance

4.1 Introduction

This chapter provides a maintenance schedule to ensure proper and safe operation after installation. Further adjustments or calibrations not contained in this "Instructions for Use", refer to the applicable "Technical Description" and have to be made by **PROTEC** personnel or expressly by **PROTEC** authorized service personnel.

Before operating this X-ray unit, user have to make themselves acquaint with all control elements and their functions.

The preventative maintenance procedures required to ensure the operational integrity and safety of the equipment are listed in the following paragraphs. It is the owner/user's responsibility to perform preventative maintenance at the specified intervals or arrange for such service with an authorized service representative.

Maintenance has to be recorded.

4.1.1 Operating

1. Careless movements of the whole table or the table top may can hit the image receptor and cause damages to it.
2. Watch for obstacles (persons, chairs, tables, carriages etc.) inside the obvious movement area of the table. Disregard can cause injuries to patient or damages to the X-ray unit or the obstacles.

1. Only trained and authorized persons are allowed to operate the system.

2. For replacement use original parts only.

4.2 Safety Information

The user and the personal have to follow the warnings and safety information's, placed on the device, disregard may lead to injuries.

The operator must make himself familiar with all warnings, placed on the device.

This is necessary for the safety and ensure the correct operation.

In the event of a malfunction do not an longer the unit and notify **PROTEC** service or the expressly authorized service provider.

4.3 Technical Safety Information

To protect the safety of patients, users and third parties, it is absolutely necessary that Checks, which ensure the reliable function and operational safety are made in intervals of 24 months by **PROTEC** service or expressly authorized service providers.

All parts of this equipment that could create a hazard through wear and tear must be checked in at least intervals of 24 months and if necessary, replaced by **PROTEC** service or by expressly authorized service providers at regular intervals.

As the manufacturer, we are responsible for the technical safety aspects of the equipment if any maintenance, repairs and/or modifications are performed by our staff or by expressly authorized service providers. Likewise, if component parts that affect the safety of the equipment are replaced by original replacement parts in the event of a failure, we are responsible for the technical safety aspects of the equipment.

In the event that scheduled maintenance is not performed, **PROTEC** will not be responsible for damages incurred by the user or third parties if such damages are the result of improper or omitted maintenance.

Before the start of examination operations, the operator must ensure that all the equipment, which is listed in the "Instructions for Use" and which is relevant to safety, is functioning properly and that the product is ready for use. A visual check shall be made to insure that all displays and indicator lamps are functioning correctly.

4.4 Technical safety checks

The technical safety checks have to be carried out all 24 month by PROTEC Service or authorized service providers to ensure the safe and reliable operation of the equipment.

In the event that scheduled maintenance is not performed, PROTEC will not be responsible for damages incurred by the user or third parties if such damages are the result of improper or omitted maintenance.

For maintenance check list refer to the manual „Technical Description“.

4.5 Maintenance schedule

4.5.1 User's maintenance schedule

Prior to cleaning or disinfecting, ensure that no liquids can penetrate into the equipment.

4.5.1.1 Cleaning

Do not use water for cleaning.

Water causes short circuits in the electrical portion of the system, and corrosion in the mechanical components.

Do not use corrosives, solvents or abrasive cleaning materials.

Clean painted and plastic surfaces only with a cloth and common household cleaners and wipe surfaces with a clean, dry, lint-free cloth. Clean chrome-plated parts by wiping with a clean, dry, lint-free cloth.

4.5.1.2 Disinfection

All components, including accessories and connecting cables, may be disinfected only by wiping with a disinfectant solution. Please contact **PROTEC** if detailed information is needed concerning appropriate disinfection materials.

The equipment must be well covered with plastic sheets during room disinfection.

For safety reasons, no spray disinfectant may be used.

Recommendations and instructions regarding the use of disinfectants and methods of disinfection may be derived from the most recent rules and guidelines concerning disinfection and explosion safety.

4.5.2 User's daily maintenance before and during operation

Check that wheels brakes are in function. (Fig. 2-1/1)

Check fixing and easy movement of tabletop by using the brake foot-pedal bar (Fig. 2-1/2).

4.5.3 Monthly Checks

4.5.3.1 Quality Control

X-ray equipment should be quality-controlled at regular intervals, at least monthly, or as required by applicable regulations to determine that the image quality remains in accordance to national regulations, e.g. by a monthly consistency testing.

4.5.4 Maintenance

Required maintenance must be performed at 24-months intervals by **PROTEC** Service or authorized service providers to ensure the safe and reliable operation of the equipment.

In the event that scheduled maintenance is not performed, PROTEC will not be responsible for damages incurred by the user or third parties if such damages are the result of improper or omitted maintenance.

4.5.5 Duration of Lifetime of the Product

The **PROTEC** PROGNOT XP is designed for a useful life of ten years, when used as intended and the regular maintenance schedule will be performed by the **PROTEC** service organization or by expressly authorized service providers.

4.5.6 Disposal Remarks

The PROGNOT XP does not contain any toxicological materials.


All mechanical, electrical and plastic components have to be disposed according to local or national regulations. In case of doubt contact **PROTEC**.

5 Combination with other equipment

- The PROGNOT XP is normally prepared for the use with X-ray units as e.g. L/U-arms for general radiography in diagnostic human medicine use.
- For combination with other equipment please contact **PROTEC** for compatibility tests and release.

6 Technical Data

6.1 Measurements



Tabletop:

200 cm x 75,5 cm, or
226 cm x 75,5 cm, or
200 cm x 65,5 cm, or
226 cm x 65,5 cm

Max. patient weight

230 kg (Composite fiber TP)
250 kg (Carbon TP)

Total weight without patient

98 kg

Height:

760 mm

Tabletop transverse movement from center pos.

± 100 mm

Tabletop longitudinal movement from center pos.

± 330 mm (at 200 cm table top)

± 460 mm (at 226 cm table top)

Distance between stroke columns below tabletop:

1119 mm

Brakes of tabletop are mechanically operated.

6.2 Attenuation equivalent of Tabletop

The aluminum attenuation of the tabletop is typical 0,7 and <0,8 mm AL according to EN 60601-1-3 at 100kV and a first half value of 3,7 mm Al and typical 0,6 mm AL and <0,8 mm according to 21CFR § 1020-30 (n) at 100 kV and a first half value layer of 2,7 mm Al.

6.3 Product Life Time

The PROGNOST XP is designed for a useful lifetime of ten years if used according to specifications and regular maintenance through **PROTEC** service or authorized service providers.

6.4 Environmental Conditions

6.4.1 Operating Environment

Temperature range

+ 10°C to + 40°C

Relative humidity range

30% to 70% (not condensing)

Atmospheric pressure range

700hPa to 1060hPa

6.4.2 Transport and stock environment

Temperature range


- 10°C to + 70°C

Relative humidity range

10% to 95% (not condensing)


Atmospheric pressure range

500hPa to 1060hPa



Label on head and foot end of tabletop:

Tabletop made of carbon



Tabletop Composite fiber 200cm



Tabletop Composite fiber 226cm



Labels on table top



230kg
506lb

Maximum allowable weight patients (line load) on table top
Composite fiber tabletop



250kg
550lb

Maximum allowable weight patients (line load) on tabletop
Carbon table top

Label on front side frame

FDA Label



Date of Manufacture:
Place of Manufacture: Oberstenfeld, Germany
This product complies with CDRH 21CFR, Subchapter J, as of the date of manufacture.
PROTEC GMBH & Co. KG

7.3 Abbreviations

mm	millimeter
cm	centimeter
kg	kilogram
°C	Degree centigrade
hPa	Hektopascal
DIN	German Industrial Standard
EN	European Norm
CE	CE-marking